

Template for comments and secretariat observations

1	2	(3)	4	5	(6)	(7)
MB¹	Clause No./ Subclause No./ Annex (e.g. 3.1)	Paragraph/ Figure/Table/ Note (e.g. Table 1)	Type of com- ment²	Comment (justification for change) by the MB	Proposed change by the MB	Secretariat observations on each comment submitted
JP	whole		ed	"ml" and "mL" are mingled in the text. Probably "mL" would be more appropriate in terms of clarity.		
JP	whole		te	The use of the symbol % is acceptable because it is internationally recognized (ISO 31-0). However, the symbol % represents simply a number. The use of phrases such as "% (v/v)" and "% (m/v)" should be avoided.	The phrase "1 % (m/m)", for example, should be replaced with "0.01 kg kg ⁻¹ " or with "0.01 kg/kg"	
JP	7.4	NOTE	te	We do agree that the Zeeman background correction system is superior to the deuterium system in ETAAS. However, the AAS apparatuses with only the deuterium system are still on the market and many people are actually using them. Some advice to overcome the difficulty in deuterium system should be added in this note or in the text.		

1 **MB** = Member body (enter the ISO 3166 two-letter country code, e.g. CN for China; comments from the ISO/CS editing unit are identified by **)

2 **Type of comment:** **ge** = general **te** = technical **ed** = editorial

NOTE Columns 1, 2, 4, 5 are compulsory.